

Certificate of Analysis



Client:
Skye Peptides

Laboratory:
TrustPointe Analytics LLC
1743 142nd Ave Suite 4
Dorr, MI 49323



Verify at TrustPointeLIMS.com
Key=KBFM8ZZF1D5B

Sample Information

Sample Name KLOW 80mg, 1 vial, ENDO **Client SID** KLOW25-80-005 **Published** 2025-12-30 09:55

Results

Sample IDs	SPL-2712
USP <85> Bacterial Endotoxin	<10.0 EU/mL
USP <85> Sample CV %	0.0 %
USP <85> Spike CV %	8.1 %
USP <85> Spike Recovery	62 %

Comments for SPL-2712

General

Original dilution 2mL; USP limit 5 EU/kg/hr

Approvals

Created by

Ashlee Lust
President

Reviewed by

Joshua Lust
Managing Director